

## Cement selection made easy.

	RelyX™ Ultimate Adhesive Resin Cement	3M™ RelyX™ Unicem 3M™ RelyX™ Unicem 2 Self-Adhesive Resin Cement	RelyX™ Luting Plus Resin Modified Glass Ionomer Cement	RelyX™ Veneer Cement
<b>Metal/Metal Based</b>				
Inlays/Onlays	+	++	+	-
Crowns/Bridges	+	++	++	-
Endodontic Posts	+	++	++	-
Maryland Bridges	++	+	-	-
On Implant Abutments	+	++	++	-
<b>Glass Ceramics (incl. Li. Disilicate) (e.g. e.max®, VITA Mark II, IPS Empress 2)</b>				
Inlays/Onlays/Table Tops	++	+	-	-
Crowns/Bridges	++	++	-	-
Veneers	+	-	-	++
On Implant Abutments	+	+	-	-
<b>Oxide Ceramics (e.g. Lava™ Plus, Brux Zir®, Procera®)</b>				
Inlays/Onlays	++	++	+	-
Crowns/Bridges	+	++	+	-
Endodontic Posts	+	++	+	-
Maryland Bridges	++	+	-	-
On Implant Abutments	+	++	++	-
<b>Resin Nano Ceramics (e.g. Lava™ Ultimate CAD/CAM Restorative)</b>				
Inlays/Onlays	++	-	-	-
Veneers	+	-	-	++
<b>Resin Composites</b>				
Inlays/Onlays	++	+	-	-
Crowns	++	++	-	-
Endodontic Posts	+	++	-	-
Veneers	+	-	-	++
On Implant Abutments	++	++	-	-
<b>++ Highly recommended*   + Recommended   - Not indicated</b> * Either better performance for this indication or easier handling with equal performance.				

**Note:** The cement selection recommendations in this guide list the key indications for a particular cement, and are intended as a general guide only. 3M RelyX cements are highly versatile products that can be used for multiple indications. Your selection is dependent on individual patient circumstances and your preferred procedural techniques.



3M Oral Care  
2510 Conway Avenue  
St. Paul, MN 55144-1000 USA

Phone 1-800-634-2249  
Web [3M.com/dentalcements](http://3m.com/dentalcements)

3M Canada  
Post Office Box 5757  
London, Ontario N6A 4T1  
Canada

Phone 1-888-363-3685

3M, "3M Science. Applied to Life.", Lava, RelyX and Scotchbond are trademarks of 3M Company or 3M Deutschland GmbH. Used under license in Canada. All other trademarks are owned by their respective owners.

Please recycle. Printed in U.S.A. © 3M 2016.  
All rights reserved. 70-2013-0776-9