



PRODUCT CLINICAL DATA SUMMARY
NO. 1567 [formerly MSX 6471]
Effective: March 2009

The adhesive used on No.1567, as part of another construction, has been subjected to the following safety evaluations:

In Vitro Cytotoxicity (L929 Agar Overlay)

Protocol reference: Guess, W. L. et al; "Agar Diffusion Method for Toxicity Screening of Plastics on Cultured Cell Monolayers" J. Pharm. Sci. 54:1545-1547 (1965).(3M study # T 7814.2)

This study meets the GLP requirements of EPA 40 CFR part 792 and 160, FDA CFR 58 as specified in The Testing of Chemicals, published by the Organization for Economic Cooperation & Development (OECD), 1997.

Results: 0.0/0.0

Acute Primary Skin Irritation in Albino Rabbits

Protocol reference: Draize: Appraisal of the Safety of Chemicals in Food, Drugs and Cosmetics (1965). Published by the Editorial Committee of the Association of Food and Drug Officials of the United States. (3M study # T 7814.1) This study was conducted in accordance with the Good Laboratory Practice requirements of the EPA, 40 CFR 160 and 792, FDA 21 CFR 58 , and the OECD, The Testing of Chemicals, 1997.

Results: 0.0/8.0

These tests are in accordance with the ISO 10993 Part-1 "Biological Evaluation of Medical Devices", as put forth by the FDA. All laboratory testing was conducted in accordance with the FDA Good Laboratory Practices Regulation of 1978.

It is the responsibility of our customers to determine the final suitability of our products for their application.