

**Recall Notice**  
**Selected Lots of**  
**3M™ Versaflo™ TR-600 Series Cartridges**  
**IMMEDIATE ACTION REQUIRED**

September 2, 2016

3M has determined that selected production lots of the following three 3M™ Versaflo™ TR-600 Series cartridges may have been vacuum sealed too tightly, which may affect the performance of the center gasket due to deformation.

The products under recall were produced between May 30, 2016 and August 4, 2016.

Product	Affected Lots	3M ID	UPC Number
TR-6510N Organic Vapor/HEPA Cartridge	0616	70-0716-1830-3	50051131373612
TR-6530N Organic Vapor/Acid Gas/HEPA Cartridge	0616 0716 0816	70-0716-1831-1	50051131373629
TR-6820N HEPA/HF/Nuisance OV	0716	70-0716-1829-5	50051131373605

\*One TR-6530N cartridge is included in TR-600-HIK kits – please review cartridge in kits.

Please review your product inventory; if you have any of these products from the specific lot codes listed above, please contact 3M at 1-800-355-6682 for return and replacement information. See Image 1 below for location of lot code on the box. The same lot code will also be on the individual cartridge packaging and on the back of the cartridge.

Some of the cartridges in these lot numbers have been inspected, tested and re-packaged by 3M – these will have a yellow dot on the box label and also on the individual cartridge package label (see Image 2 below). Products with the yellow dot are exempt from the recall.

We have corrected the issue at our production facility and are working diligently to restore supply. Please note that this recall is only for the three products and lot numbers listed above and does not affect any other TR-300 or TR-600 Series filters, cartridges or the TR-6500FC filter covers.

We regret any inconvenience this may cause and appreciate your continued patience. Should you have any questions please contact your local Sales Representative or Technical Service at 1-800-243-4630. In Canada, please call 1-800-267-4414.

Image 1 – LOT code identifier:



Image 2 - Inspected, tested and repackaged product (note yellow dot)  
- Exempt from recall:



Sincerely,

Brandon Cordts  
3M Personal Safety Division  
Respiratory Products Manager

Karl Schroeder  
3M Personal Safety Division  
Respiratory Products Quality Manager