

3M™ Tegaderm™ PICC/CVC Securement
Device + CHG Chlorhexidine Gluconate I.V. Securement Dressing



Application & Removal Guide

Silicone
σιλικόνης
Szikon

3M

3M™ Tegaderm™ PICC/CVC Securement Device + CHG Chlorhexidine Gluconate I.V. Securement Dressing

Device, Dressing, Tape Strip Application Instructions

Minimize catheter manipulation during application.



1 Allow all preps and skin protectants to dry completely. Open package and remove securement device. Ensure you hold the insertion site with sterile tape or a sterile gloved finger. Before placing the securement device on the skin, orient the device with arrows pointing toward the insertion site.



2 Place single, double or triple lumen CVC catheter into device and weave lumen(s) under the single plastic arm.



3 Remove the liner of the attached tape strip and secure lumen(s) to the device base.



4 Position device on skin at the desired location. Pull and remove liner from one side of base to expose adhesive while holding the device in place. Pull and remove liner from the other side of base.



5 Apply pressure to the device base to establish good adhesion to the skin. Apply the 3M™ Tegaderm™ I.V. Advanced Dressing following steps 6-10.



6 Peel the liner from the dressing, exposing the adhesive surface.



7 Place the dressing so the CHG gel pad covers the insertion site and the dressing border covers the plastic base of the device. Do not stretch dressing during application. Mechanical skin trauma may result if the dressing is applied with tension.



8 Apply firm pressure on the dressing, including the edges, to enhance adhesion to the skin.



9 Slowly remove the frame while simultaneously applying manual pressure to the dressing edges.



10 Smooth the dressing from the centre toward the edges using firm, but gentle pressure to enhance adhesion.



11 Remove the liner from the sterile securement tape strip.



12 Grasp the non-adhesive tab of the securement tape strip, and bend slightly with thumb.



13 Lift the catheter lumen(s) and apply the notched end of the securement tape strip under the catheter lumen(s) and over the dressing edge. Push the tape strip notch forward, abutting up against the catheter lumen(s). Apply pressure on the securement tape strip to enhance adhesion.



14 Slowly remove the frame from the securement tape strip while smoothing down the edges.



15 Document dressing change on the documentation tape strip according to facility protocol. Place the documentation tape strip on top of the dressing over the catheter lumen(s). Apply pressure to enhance adhesion.

3M™ Tegaderm™ PICC/CVC Securement Device + CHG Chlorhexidine Gluconate I.V. Securement Dressing

Device, Dressing, Tape Strip Removal Instructions

Minimize catheter manipulation during removal.



1 Lift catheter lumens, exposing the notches of the perforated tape strip. Gently pull apart the perforated tape strip. Avoid skin trauma by peeling the dressing back, rather than pulling it up from the skin.



2 Lift the catheter lumen(s) with one hand and place gloved index finger on the base of the device. Utilizing the low and slow removal technique, slowly start removing the dressing toward the insertion site.



3 When the catheter hub is exposed, move your gloved finger to secure the catheter hub and continue to remove the dressing until the device is uncovered. Leave the remainder of the dressing in place over the catheter insertion site.



4 Remove the device tape strip from the catheter lumen(s).



5 Use a gloved finger to stabilize the catheter hub and gently remove the catheter lumen(s) out from under the plastic arm of the device.



6 Secure the catheter with one hand and use your other hand to remove the device from the patient's skin.



7 To remove the remainder of the dressing, stabilize the catheter with a tape strip or gloved finger. Grasp the corner of the CHG gel pad and the dressing between thumb and finger.



8 Apply sterile alcohol, saline or water between the gel pad and skin to facilitate removal using the low and slow technique.



▶ visit [3M.com/3MSecurement](https://www.3m.com/3MSecurement) to watch application and removal videos

Suggested Applications

- Peripherally Inserted Central Catheters (PICCs)
- Central Venous Catheters (CVCs)

Securement System Selection

Choose the system with a dressing large enough to provide at least a one-inch margin of adherence on dry, healthy skin around the catheter site.

Site Preparation

- Prepare the site according to your facility's protocol.
- Clipping hair at the site may improve dressing adhesion. Shaving is not recommended.
- 3M™ Cavilon™ No Sting Barrier Film may be used with the 3M™ PICC/CVC Securement System. Apply Cavilon™ No Sting Barrier Film evenly to the skin. Apply a sterile, alcohol-free barrier film compatible with your skin antiseptic if needed.
- Allow all preps and protectants to dry completely before applying the securement device and 3M™ Tegaderm™ I.V. Advanced Dressing to ensure good adhesion and to help prevent skin irritation.

Site Care

- Follow universal precautions as indicated by established protocols for infection prevention.
- Observe site daily for signs of infection or other complications. Infection may be signaled by fever, pain, redness, swelling, or unusual odour or discharge.
- If infection is suspected, remove the 3M™ Tegaderm™ I.V. Advanced Dressing, inspect the catheter site directly, and determine appropriate medical intervention.
- Inspect the 3M™ PICC/CVC Securement System daily and change the system as necessary, in accordance with facility protocol.
- Securement System changes should occur at least every 7 days and may be needed more frequently with highly exudative sites or if integrity of the device or dressing are compromised.

Reducing Infection Risks

- At each securement system change, pay careful attention to apply skin antiseptic to the skin around and under the catheter.
- Carefully disinfect injection ports before access.
- Inspect the site frequently for early signs of complications. Transparent dressings allow easy site assessment without removing the dressing.
- Change the securement system according to your facility's protocol, or when it becomes wet, loosened or soiled. Edge lift of a transparent dressing is not necessarily failure, unless there is a channel from the edge of the dressing to the insertion site of the I.V. catheter.
- Use maximum barrier precautions for central venous catheter (CVC) insertion and sterile technique for site care.
- Protect against skin irritation. Compromised skin near the catheter insertion site increases the risk of complications.
- No dressing can substitute for your professional site care.

Warnings

- Do not use 3M™ PICC/CVC Securement Device + Tegaderm™ I.V. Advanced Dressing ("Securement System") where loss of adhesion could occur, such as with a confused patient, diaphoretic or non-adherent skin, or when the access device is not monitored daily.
- Failure to follow the manufacturer's instructions for use may result in complications including inadequate securement and skin irritation.
- This is a single use securement system. Reuse and/or repackaging may create a risk of patient or user infection, compromise the structural integrity and/or essential material and design characteristics of the Securement System, which may lead to device failure.

Tegaderm™ CHG Dressing should be changed:

- If there is visual drainage outside the gel pad.
- If the dressing becomes loose, soiled, or compromised in any way.
- If the site is obscured or no longer visible.
- NOTE: Tegaderm™ CHG Dressing is not designed to absorb large quantities of blood or drainage.
- If the dressing appears to be saturated or overly swollen.

For a complete listing of all product warnings and precautions please view the product insert included with each product.

Ordering Information

Product Name	Product Code	Overall Device Size	Overall Dressing Size	Device + Dressing/Box	Boxes/Case
3M™ PICC/CVC Securement Device + Tegaderm™ I.V. Advanced Securement Dressing	1837-2100	2 in x 2 1/8 in 5.1 cm x 5.4 cm	3 1/2 in x 4 1/2 in 8.5 cm x 11.5 cm	20	4
	1839-2100	2 in x 2 1/8 in 5.1 cm x 5.4 cm	4 in x 6 1/8 in 10 cm x 15.5 cm	20	4
3M™ Tegaderm™ PICC/CVC Securement Device + CHG Chlorhexidine Gluconate I.V. Securement Dressing	1877R-2100	2 in x 2 1/8 in 5.1 cm x 5.4 cm	3 1/2 in x 4 1/2 in 8.5 cm x 11.5 cm	20	4
	1879R-2100	2 in x 2 1/8 in 5.1 cm x 5.4 cm	4 in x 6 1/8 in 10 cm x 15.5 cm	20	4



Available in Canada from:

Critical & Chronic Care Solutions Division
3M Canada Company
 P.O. Box 5757
 London, Ontario N6A 4T1
 Canada
 1-800-563-2921
 www.3M.ca/skinwound

3M Health Care
 2510 Conway Avenue
 St. Paul, MN 55144-1000
 USA
 1-800-228-3957
 www.3M.com/C3SD

Please recycle. Printed in Canada.
 © 2015, 3M. All rights reserved.
 3M and Tegaderm are trademarks of 3M.
 Used under license in Canada.
 1509-03504 E