

3M™ Petrifilm™ Environmental *Listeria* Plates

Product Information

MEDIUM:	3M Petrifilm Environmental <i>Listeria</i> Plate (6447/6448)										
ISO/BSI CERTIFICATE OF REGISTRATION NUMBER:	Manufactured at Wroclaw, Poland (ISO 9001:2015, FM 581890) Manufactured at Brookings, SD, USA (ISO 9001:2015, FM 14552)										
DATE OF EXPIRATION/ LOT NUMBER:	Expiry and lot number indicated on each package. Lot number indicated on each plate.										
FORMULATION:	Selective agents, nutrients, cold-water soluble gel, and a chromogenic indicator										
METHOD OF PREPARATION:	Nutrients and gels coated onto films. For use, hydrate with three mL aqueous sample or dilution of sample. See product instructions for detailed information.										
CONTAMINATION CHECK:	96 randomized plates per batch tested Incubated at 35°C for 28 hours Utilizes statistical sampling plan										
EFFICACY CHECK:	Complement of organisms tested includes, among others: <table><thead><tr><th><u>Organism</u></th><th><u>Acceptable batch</u></th></tr></thead><tbody><tr><td><i>Listeria monocytogenes</i> SLR-22</td><td>Typical Growth</td></tr><tr><td><i>Listeria innocua</i> ATCC 33090</td><td>Typical Growth</td></tr><tr><td><i>Enterococcus faecalis</i> ATCC 14506</td><td>No or Atypical Growth</td></tr><tr><td><i>Bacillus atrophaeus</i> ATCC 9372</td><td>No or Atypical Growth</td></tr></tbody></table>	<u>Organism</u>	<u>Acceptable batch</u>	<i>Listeria monocytogenes</i> SLR-22	Typical Growth	<i>Listeria innocua</i> ATCC 33090	Typical Growth	<i>Enterococcus faecalis</i> ATCC 14506	No or Atypical Growth	<i>Bacillus atrophaeus</i> ATCC 9372	No or Atypical Growth
<u>Organism</u>	<u>Acceptable batch</u>										
<i>Listeria monocytogenes</i> SLR-22	Typical Growth										
<i>Listeria innocua</i> ATCC 33090	Typical Growth										
<i>Enterococcus faecalis</i> ATCC 14506	No or Atypical Growth										
<i>Bacillus atrophaeus</i> ATCC 9372	No or Atypical Growth										
PACKAGING:	Pack size: 25 plates per foil pouch Film grade: Plastic foil laminate Seal integrity check: Pressure Decay test performed										
MEASUREMENT SYSTEMS CALIBRATION AND TRACEABILITY:	Incubator temperature 3M internal calibration Minimum calibration once per year for all equipment										
MEDIA QUALITY: STATEMENT	Certificate of Analysis available on 3m.com/foodsafety										
SHELF LIFE:	18 months from date of manufacture										
STORAGE CONDITIONS:	Store at temperatures less than or equal to 8°C										